

Care and use guide



Contents	page
Features	2
Quick guide	2
Before first use begins	3
Removing and replacing the storage cap/dibber	3
Hydration	4
MPORTANT - Soil pH Pen probe care	4
To operate	6
nformation about measuring the oH of soils/media	7
Measuring soil pH value	9
Measuring pH value of a soil solution	10
Cleaning	11
Battery replacement	11
Calibration	12
Error messages	12
Troubleshooting guide	13
Technical specifications	14
Bluelab pH probe care tools	14
Bluelab limited warranty	15
Contact details	16





Features	
Direct pH and temperature measurements from soils/media	Selectable units for temperature
	Automatic Temperature Compensation (ATC)
Backlit LCD display	Low battery indicator
Hold reading function	Fully waterproof
Integrated storage cap/dibber	Auto off function
Successful pH calibration indicator	Fully guaranteed for 1 year

Quick guide



Check mark/tick to indicate successful pH calibration

Disappears 30 days after last successful calibration as a reminder calibration is due.

Low battery warning

Appears when batteries are low.

Power button / hold

Short press to turn on. Short press to hold and unhold reading. Long press to turn off.

Calibrate button

See calibration section.

Units button

Press and hold until units flash then short press to change units. Screen will change back to main display when no buttons have been pressed for 4 seconds.

Storage cap/dibber

Create a pathway in the soil/media before taking a measurement with the pH probe. Use to store the pH probe tip, to keep it wet at all times.

ATTENTION:

Always loosen the cap!

Check for the gap. Always loosen the cap before removing or placing on the pH probe tip. Refer to section 2.0

FRAGILE Glassware

The pH probe contains a glass tube and a glass bulb. DO NOT drop, knock or bend. Refer to section 4.0

Keep your probe tip wet at all times to avoid permanent damage. Refer to section 4.0







1.0 Before first use begins


ATTENTION: Salt-crust forming on the pH probe is normal. A small amount of KCl solution from the storage cap/dibber may leak out leaving a salt-crust. This does not affect the probe and is normal.

The following tasks must be performed before the Bluelab Soil pH Pen is used for the first time.

Check the soil pH probe tip is covered with KCl storage solution

pH probes need to be be kept wet at all times. Before you remove the storage cap/dibber, check how much Bluelab pH Probe KCI Storage Solution is in the cap.

- a) Stand the soil pH pen upright. There should be enough KCl storage solution to cover the pH probe tip. If the probe tip is covered, go to steps 2 & 3.
- b) If the probe tip is NOT submerged in KCI storage solution, you will need to hydrate the soil pH probe before use. Remove the storage cap (see secton 2.0), then go to section 3.0 for hydration steps.




- Removing and replacing the storage cap/dibber IMPORTANT: Pulling the cap off, or placing the cap on, without loosening first will break the probe. See section 2.0 for more details.
- Calibrate the Soil pH Pen before the first use to ensure accurate readings. Go to section 11.0 for calibration steps.

2.0 Removing and replacing the storage cap/dibber

IMPORTANT: Always loosen the storage cap/dibber before removing or replacing from the probe.

- To remove the storage cap/dibber
 - a) Grip the top and twist the dibber a few times to the left to loosen. You'll know the cap is loose when you can see a gap between the cap and the dibber.
 - b) Slide the storage cap/dibber off.
 - Place cap upright, in a cup or similar container so that the KCl storage solution doesn't spill out.
- To replace the storage cap/dibber Place the loosened storage cap/dibber gently on the probe, sliding until it stops moving up. Grip the top and twist the cap to the right

until secure.



Hydration 3.0

Hydrate the soil pH pen in Bluelab pH Probe KCl Storage Solution when:


- the probe tip has not always been stored in KCl storage solution, to improve the reading response speed.
- the probe tip has been accidentally allowed to dry out

Never use RO, Deionized or Distilled water. Pure water changes the chemistry in the reference, causing the probe to die.

- Loosen, then remove the storage cap. Place the soil pH pen upright in a plastic container.
- Clean the pH probe tip. Ensure the probe tip is cleaned before hydrating. See section 9.0 for instructions.

Cleaning before the first use is not required.

- Add enough Bluelab pH Probe KCI Storage Solution to submerge the probe tip.
- Leave to soak for at least 24 hours. After hydration, always calibrate the soil pH pen to ensure accuracy, see section 11.0.



4.0 IMPORTANT - Soil pH Pen probe care

To ensure you receive a long life from your Soil pH Pen, please ensure you follow the guide below.

Soil pH Pen body & probe Storage cap/dibber Pen body Cap Seal **Temperature** probe and internal Soil level Dibber glass tube - fragile indicator Bluelab pH Glass bulb Probe KCI fragile Storage Solution

The Soil pH Pen probe is fragile

- DO NOT drop, knock or exert a sideways force on the probe.
- · Avoid sudden temperature changes.

Note: this section continues on next page...



4.0 IMPORTANT - Soil pH Pen probe care cont.

The soil pH probe does not last forever

- Probes do not last forever. They age through normal use and will eventually fail.
 The lifetime of a probe depends on the environment it is used in and the way it is treated.
- Always loosen the cap before removing or replacing the storage cap/dibber.
- This probe is suitable to use in temperatures between 0 50 °C / 32 122 °F only.
- Chemically aggressive, abrasive or oily samples will reduce the life of the probe.

The Soil pH Pen probe requires cleaning at least once a month and prior to calibrating

Cleaning the glassware is critical to obtaining accurate measurements.

The calibration check mark will disappear from the soil pH pen screen 30 days after the last successful calibration to remind you it is time to clean and calibrate again.

Storing the Soil pH Pen

The probe tip must be kept wet - if it dries, it dies!

To prepare the probe for storage, place enough Bluelab pH Probe KCl Storage Solution, to cover the probe tip, into the storage cap/dibber. Never use RO, distilled or deionized water. Place the loosened cap over the probe tip and tighten (see section 2.0).

Long term storage

For long term storage, sit the Soil pH Pen upright, ensuring the storage cap/dibber contains enough Bluelab pH Probe KCl Storage Solution to cover the probe tip. Check each month to ensure the liquid has not evaporated.

If the probe has been accidentally allowed to dry out:

The probe must be 'hydrated' by soaking for 24 hours in KCl storage solution, refer to section 3.0 for details. Following this; carry out a calibration to check if the probe has already suffered permanent damage, see section 11.0.

DO NOT let the probe tip dry.

DO NOT exert sideways force, drop or knock the pen. This will break its external glass bulb or internal glass tube.

DO NOT touch the glass bulb with your fingers as this will contaminate the glass.

DO NOT plunge a cold probe into a hot liquid (or vice versa). Sudden temperature changes can crack the glass and permanently damage the pen.

DO NOT immerse in oils, proteins or suspended solids that will leave a coating on the glass bulb.

DO NOT remove or place the storage cap/dibber back onto the probe without first loosening the cap, as you could reduce the life of the probe.

DO NOT store, soak or rinse the pH probe in RO (Reverse Osmosis), Distilled or Deionized water. Pure water changes the chemistry in the reference, causing the probe to die.



5.0 To operate

1 Turn Soil pH Pen on

Press power button. The last measurement is recalled for 3 seconds.

To turn Soil pH Pen off

Press and hold the power button until OFF is displayed.

NOTE: The Soil pH Pen will automatically turn off after 4 minutes to conserve battery power.

2 Measure pH

- a) With the storage cap/dibber on the pen, press the Soil pH Pen into the sample area until it reaches the word 'soil' at the top of the dibber. Remove gently from soil or substrate, leaving a small hole.
- b) Remove storage cap/dibber (refer to section 2.0 for instructions). Turn the Soil pH Pen on, then place pH probe into the newly made hole in the soil or substrate. Do not exert sideways pressure. Wait for the reading to reach a constant value.

NOTE: Never make a hole in the soil or substrate with the pH probe. *Always* use the storage cap/dibber.

To hold reading

If you want to "hold" the reading on the screen, short press the power button. To exit the hold function, short press the power button again.



1 second alternating displays

To change temperature units

Hold down the units button for 3 seconds until the temperature units start flashing. Release. Short press the units button again to cycle between "F and "C. To exit this mode don't press anything for 3 seconds.


NOTE: You can change units while in hold mode.


Skinse and place on the storage cap/dibber

To ensure accurate pH readings always rinse the probe in clean fresh tap water before placing on the storage cap/dibber. Ensure there is enough Bluelab pH Probe Storage Solution in the storage cap/dibber to cover the probe tip. To place the storage cap/dibber back on, see section 2.0.



Power button









6.0 Information about measuring the pH of soils/media

pH is the measurement of the hydrogen ion concentration (H+) - acidity and its opposite, alkalinity. Neutral pH is 7.0 pH. Acidity measures below seven pH (7.0 pH) with alkalinity measuring above it (7.0 pH). See chart below.


In soils or growing media, pH strongly influences the availability of nutrients and the presence of microorganisms in the soil.

Certain plants require a particular pH range to enable the required nutrients to be consistently available to the plant. If the solution is too acidic or too alkaline it can cause "lock up" - a situation which restricts certain elements essential for growth from being absorbed by the root structure. This in turn reduces plant health and performance. Deficiencies in the required elements become apparent in plant growth and can lead to crop failure.

Low soil pH causes aluminium and manganese toxicity in plants and reduces the availability of soil phosphorus. High soil pH also reduces soil phosphorus availability and reduces micro nutrients such as zinc and boron to plants.

The chart below shows how nutrient pH levels influence the uptake of certain elements.

Recommended pH range Solution 5.5 - 6.3 Soil 6.2 - 7.2 for plants arown in: can be plant specific



6.0 Information about measuring the pH of soils/media cont.

Taking pH measurements of soils with an electronic meter is indicative rather than absolute.

The following factors are outside the control of any soil pH meter, so to minimise their effect on the accuracy of the pH measurement you should consider the following precautions:

Moisture level/raw water

If the sample you are wishing to measure is dry, add RO water or distilled water to moisten. Ideally wait 24 hours before you take a measurement.

NOTE: If you add tap water, you will influence the soil pH reading based on the pH of the tap water.

Calibration of the Soil pH Pen and cleanliness of the soil probe tip

Calibrating the Soil pH Pen at least monthly will ensure accurate readings. Cleaning the soil residue from the probe tip and storing the Soil pH Pen in a clean moist state will provide reliable readings as well as prolonging the pens life.

Sample selection

For field testing, remove the top $5-10\ cm/2-4"$ of the top of the soil. Samples are taken approximately $15-20\ cm/6-8"$ down into the substrate and from various areas, then an average of the readings is used.

For container grown plants, it is recommended to check the pH level of the substrate prior to planting.

Factors affecting pH in the soil or media:

Soil type

Soils formed under high rainfall conditions (e.g. Eastern USA) are more acidic than those formed under dry conditions (e.g. Western USA).

Growth stage of the plant

Uptake and requirements of particular elements change as the plant progresses through it's growing cycle. Recording pH level data to create a history is valuable.

Applications and types of fertilizers

Applications and types of fertilizers can alter the pH level significantly. The time at which you take the reading is important. Evaluate the brand of fertilizer to see if it is altering the pH in the wrong direction.

Applications of sprays

As sprays can soak into the soil/media, a change to the pH level could result.

Soil/media temperature

High temperature soils may have a high concentration of CO₂. The higher the concentration of carbon dioxide pressure results in more carbonic acid which lowers pH.

pH range for soil crops

The recommended pH range for soil crops is 6.2 - 7.2, but this is plant specific.



7.0 Measuring soil pH value

Follow the steps below to take pH measurements of soil and media.

- Remove the top layer of soil/media from the surface of the sample area.
- Insert the storage cap/dibber into the sample up to the 'soil' indicator as shown on the storage cap/ dibber, then remove.

TIP: The dibber creates a safe pathway for the soil probe, reducing the likelihood of probe breakage. It should be used at all times.

Remove the storage cap/dibber from the pen and turn the Soil pH Pen on.

TIP: to keep the solution or tap water in the storage cap/dibber while taking measurements with the Soil pH Pen, simply place the storage cap/dibber upright into the soil.


- Insert the probe into the hole made by the dibber, ensuring the probe end makes gentle contact with the soil.
- Wait for the reading displayed to stabilize to a constant value. Record the reading.


TIP: If the Soil pH Pen turns off while taking a measurement, simply press the power button to turn the Soil pH Pen back on and continue with your measurement.

Remove the probe gently from the soil/media and rinse the probe tip under fresh running water (not RO or distilled) to remove any soil residue.

IMPORTANT: Always rinse the probe tip in between each measurement and shake to remove excess water.

For field testing, repeat the procedure in different locations and take the average of the measured data as the pH level is representative of the sample area.







8.0 Measuring the pH value of a soil solution

The greatest source of error in soil analysis comes during sample collection. An effort should be made to ensure each sample properly represents the area being sampled.

Collection of sample

- Sample in a zig-zag pattern across the required area.
- 2 Take samples of soil at approximately 20 cm / 8" deep.
- 3 Mix all collected samples together thoroughly.
- 4 Ideally, allow to dry in the air or in an oven at 40 °C / 104 °F.
- 5 Weigh out 20 g / 0.7 oz of the collected soil into a 150 ml / 5 fl oz plastic sample jar.

Sample preparation

- 4 Add 100 ml / 3 fl oz of distilled or deionized water, screw lid on tightly.
- Shake continuously for 5 minutes. Leave overnight and shake again the next morning.
- 3 Allow to settle for 15 minutes after shaking and strain sample into clean measuring cup.

Take pH readings as follows:

- Remove the storage cap/dibber and insert the probe tip into the soil solution sample.
- Turn the Soil pH Pen on.
- Wait for the reading displayed to stabilize to a constant value. Record the reading.

If the Soil pH Pen turns off while taking a measurement, simply press the power button to turn the Soil pH Pen back on and continue with your measurement.

- Remove the probe from the soil solution and rinse the probe tip under fresh running water (not RO or distilled) to remove soil residue.
- Place the storage cap/dibber back onto the probe after use, ensuring it contains Bluelab pH Probe Storage Solution or plenty of tap water.
 - "IF IT DRIES IT DIES!". See section 4.0 Soil pH Probe Care.

9.0 Cleaning

To ensure accurate readings the probe tip needs to be kept clean at all times. Cleaning prior to calibration is necessary for successful calibration.

- Remove storage cap/dibber. Rinse probe tip under fresh tap water.
- Fill small plastic container with clean water. Add a small amount of Bluelab pH Probe Cleaner or mild detergent (dishwashing liquid).
- Gently stir the probe tip in the mixture. Ensure that you do not 'knock' the probe on the side of the container as this may cause damage to the probe. Rinse well under fresh running water to remove all traces of the detergent mixture.
- 4 If the probe requires removal of heavy contamination: Gently brush around the glassware with a few drops of Bluelab pH Probe Cleaner or mild detergent (dishwashing liquid) and a soft toothbrush.
- Rinse well under fresh running water to remove all traces of the detergent mixture.
- Calibration of the probe is required after every clean. See the pH calibration in section 11.0. Place storage cap/dibber back onto the probe tip.



10.0 **Battery replacement**

The pH pen is powered with 1 x AAA alkaline battery. Do not use rechargeable batteries. A low battery warning is indicated by a battery symbol appearing on the screen. Only remove the battery cap when the batteries require changing. Battery life is expected to be approx. 350 hours.

- To remove old battery
 - Undo battery cap fasteners. Remove battery cap and tip out the old battery.
- Fit new battery

Insert the new battery positive (+) end down into the body.

Ensure waterproof battery cap seal is clean, and free from any debris.

Seal will fail if any dirt or debris is present.

Replace battery cap

Tighten fasteners on battery cap. Do not over tighten. Ensure silicone seal on battery cap is fully enclosed in pen body. This ensures the unit remains 100% waterproof.



Waterproof seal



11.0 Calibration

pH calibration is required before first use to ensure that the first reading is accurate. Calibration is also required when:

- The check mark/tick has disappeared from the LCD screen (30 days after last successful calibration)
- The reading is different from what you expected
- After cleaning
- After changing the battery


pH 7.0 and pH 4.0 solutions are required for calibration and should be decantered into small, clean plastic containers. You may also calibrate using pH 7.0 and pH 10.0 solutions if your readings will normally be higher than 7.0 pH.


- Excluding first use, YOU MUST CLEAN the probe tip before calibrating. See section 9.0 Cleaning.
- Remove storage cap/dibber. Refer to section 2.0, step 1.
- Rinse probe tip in fresh water and place in pH 7.0 solution.

Long press the cal button until display flashes.

- Release button. leave probe in solution while display flashing. When CAL 7 is displayed, 1 point calibration is achieved.
- 6 Rinse probe in fresh water and place it in either pH 4.0 or pH 10.0 solution (use pH 10.0 solution if you expect to measure above 7.0 pH).
- Press the cal button until display flashes. Release button. leave probe in solution while display flashing. CAL 4 or CAL 10 should be displayed (depending on what solution you are calibrating in). The check mark/tick is displayed when a 2 point (or 3) calibration is completed.

NOTE: For a three point calibration repeat steps 2, 3 & 4 using pH 7.0, 4.0 then 10.0 solution.





12.0 **Error messages**

The following error messages appear for the following reasons.

LI- °F рН

Temperature under range



Temperature over range



pH calibration failed



pH over range



pH under range



Hardware error



13.0 Troubleshooting guide				
Trouble	Reason	Correction		
	Glassware not clean	Clean glassware and calibrate		
Drift - readings slowly varying	Wick contaminated, blocked or dry	Soak probe in Bluelab pH Probe Storage Solution or tap water only for 24 hours and retest. Do not measure proteins or oils with this unit. Replacement of unit may be required		
	Glassware aged	Replace unit		
Displays similar pH reading in all buffers no matter what the buffer value is	Glassware cracked or broken	Replace unit		
Unsuccessful pH calibration	Buffers inaccurate	Replace buffers		
	Glassware not clean	Clean glassware		
	Glassware aged (glassware will not clean)	Replace unit		
	Probe not hydrated	Soak probe in Bluelab pH Probe Storage Solution or tap water only for 24 hours and retest. Do not measure proteins or oils with this unit. Replacement of unit may be required		
Noisy - readings jumping	Contact zone not immersed	Ensure soil/substrate is damp		
	Probe not adequately hydrated	Soak probe in Bluelab pH Probe Storage Solution or tap water only for 24 hours and retest. Do not measure proteins or oils with this unit. Replacement of unit may be required		
Displays pH 7 for all buffers	Glassware broken	Replace unit		
Incorrect sample reading following successful pH calibration	Wick blocked	Soak probe in Bluelab pH Probe Storage Solution or tap water only for 24 hours and retest. Do not measure proteins or oils with this unit. Replacement of unit may be required		

V	III <u>/</u> /
A	N

14.0 Technical specifications			
Measurement range	0.0 - 14.0 pH, 0 - 50 °C, 32 - 122 °F		
Resolution	0.1 pH, 1 °C/1 °F		
Accuracy at 25 °C / 77 °F	± 0.1 pH, ± 1 °C/±2 °F		
Temperature compensation	Automatic		
Operating temperature	0 - 50 °C, 32 - 122 °F		
Calibration	Two point, pH 7.0 and pH 4.0 or pH 10.0		
Units	pH, °F and °C		
Power source	1 x AAA alkaline battery		

Bluelab Probe Care - pH

The instrument is only as accurate as the probe is clean!

Probe cleaning is one of the most important parts of owning and operating any Bluelab meter, monitor or controller.

If the probe is contaminated (dirty) it affects the accuracy the reading displayed. Cleaning the pH probe is a very easy task and will prolong the life of the probe.



Bluelab Probe Care Kit - pH contains:

- Probe care instructions
- 3 x plastic cups
- 20ml single-use Bluelab Solution Sachets, 2 each of: pH 7.0 & pH 4.0, KCl
- Bluelab pH Probe Cleaner
- Toothbrush (pH probe cleaning instrument)

Bluelab pH Probe KCI Storage Solution

The perfect solution to store and hydrate your Bluelab pH products.

Bluelab pH Probe KCI Storage Solution is designed to increase response time and maximize the life of Bluelab pH pens and pH probes.

For best results, use the KCl solution to store the pH pen/ probe after use and hydrate monthly.

Instructions are on the label of the bottle.



Use Bluelab pH Probe KCI Storage Solution with:

Bluelab pH Pen

Bluelab pH Probes

Bluelab Soil pH Pen

Bluelab Soil pH Probes



Bluelab® limited warranty

Bluelab® Corporation Limited (Bluelab) provides a warranty on its products (Bluelab® Soil pH Pen) under the following terms and conditions:



How Long Does Coverage Last?

Bluelab® warrants the Bluelab® Soil pH Pen (Product) for a period of 12-months from date of purchase by original purchaser or consumer. Proof of purchase, to Bluelab's sole satisfaction, is required for the warranty to be effective (store sales receipt for Product showing model number, payment and date of purchase). This warranty is non-transferable and terminates if the original purchaser/consumer sells or transfers the Product a third party.

What is Covered?

Bluelab® warrants the Product against defects in material and workmanship when used in a normal manner, in accordance with Bluelab® instruction manuals. If Bluelab® is provided with valid proof of purchase (as defined above) and determines the Product is defective, Bluelab® may, in its sole discretion either (a) repair the Product with new or refurbished parts, or (b) replace the Product with a new or refurbished Product.

Any part or Product that is replaced by Bluelab® shall become its property. Further, if a replacement part or Product is no longer available or is no longer being manufactured, Bluelab® may at its sole discretion replace it with a functionally-equivalent replacement part or product, as an accommodation in full satisfaction of the warranty.

What is NOT covered?

This warranty does not apply to equipment, component or part that was not manufactured or sold by Bluelab®, and shall be void if any such item is installed on a Product. Further, this warranty does not apply to replacement of items subject to normal use, wear and tear and expressly excludes:

- · Cosmetic damage such as stains, scratches and dents
- Damage due to accident, improper use, negligence, neglect and careless operation or handling of Product not in accordance with Bluelab® instruction manuals, or failure to maintain or care for Product as recommended by Bluelab®
- Damage caused by use of parts not assembled/installed as per Bluelab® instructions
- · Damage caused by use of parts or accessories not produced or recommended by Bluelab®
- · Damage due to transportation or shipment of Product
- Product repaired or altered by parties other than Bluelab® or its authorised agents
- · Product with defaced, missing or illegible serial numbers
- Products not purchased from Bluelab® or a Bluelab®-authorised distributor or reseller.

How Do You Get Service?

To begin a warranty claim you must return the Product to the point of purchase with valid proof of purchase (as defined above). In California, you can also return the Product to any Bluelab-authorised distributor or reseller, with valid proof of purchase.

Limitation of Liability & Acknowledgments

TO THE MAXIMUM EXTENT PERMITTED BY LAW, THIS WARRANTY AND THE REMEDIES SET OUT ABOVE ARE EXCLUSIVE AND IN LIEU OF ALL OTHER WARRANTIES, GUARANTEES AND REMEDIES (ORAL OR WRITTEN, EXPRESS OR IMPLIED).

EXCEPT AS PROVIDED IN THIS WARRANTY AND TO THE MAXIMUM EXTENT PERMITTED BY LAW, BLUELAB IS NOT RESPONSIBLE FOR SPECIAL, INCIDENTIAL OR CONSEQUENTIAL LOSS OR DAMAGES, OR ANY OTHER LOSS OR DAMAGES RESULTING FROM SALE OR USE OF THE PRODUCT, OR BREACH OF WARRANTY, HOWEVER CAUSED, INCLUDING DAMAGES FOR LOST PROFITS, PERSONAL INJURY OR PROPERTY DAMAGE.

IT IS UNDERSTOOD AND AGREED BY CONSUMER UPON PURCHASE OF A PRODUCT THAT, EXCEPT AS STATED IN THIS WARRANTY, BLUELAB IS NOT MAKING AND HAS NOT MADE ANY EXPRESS OR IMPLIED WARRANTY OR OTHER REPRESENTATION REGARDING THE PRODUCT, AND DISCLAIMS ANY WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE TO THE EXTENT PERMITTED BY LAW. ANY WARRANTIES WHICH ARE IMPOSED BY LAW AND CANNOT BE DISCLAIMED ARE HEREBY LIMITED IN DURATION TO THE PERIOD AND REMEDIES PROVIDED IN THIS WARRANTY.

SOME JURISDICTIONS (STATES OR COUNTRIES) DO NOT ALLOW EXCLUSION OR LIMITATION FOR INCIDENTIAL OR CONSEQUENTIAL DAMAGES, OR LIMITATION ON HOW LONG AN IMPLIED WARRANTY LASTS, SO THE ABOVE LIMITATION OR EXCLUSION MAY NOT BE APPLICABLE.

IF ANY PROVISION OF THIS WARRANTY IS JUDGED TO BE ILLEGAL, INVALID OR UNENFORCEABLE, THE REMAINING PROVISIONS OF THE WARRANTY SHALL REMAIN IN FULL FORCE AND EFFECT.

Governing Law; Authority

This warranty is governed by the laws of the state of country where Product is purchased, without regard to its choice of law principles. Except as allowed by law, Bluelab does not limit or exclude other rights a consumer may have with regard to the Product. No Bluelab distributor, employee or agent is authorised to modify, extend or otherwise change the terms of this warranty.

Register your guarantee online at bluelab.com





guarantee.

The Bluelab® Soil pH Pen comes with a 1 year limited written guarantee. Proof of purchase required.



lets talk.

If you need assistance or advice - we're here to help you.

North America Ph: 909 599 1940

NZ Ph: +64 7 578 0849

Fax: +64 7 578 0847

Email: support@bluelab.com



get online.

Looking for specifications or technical advice?
Visit us online at **bluelab.com** or **facebook.com/getbluelab**



To watch instruction videos, visit our online video library: vimeo.com/bluelab



post.

Bluelab[®] Corporation Limited

8 Whiore Avenue, Tauriko Business Estate Tauranga 3110, New Zealand



Instruction Manual English PENSOILPH_V02_210916
© Copyright 2011, all rights reserved, Bluelab® Corporation Limited