



Certificate of Compliance

Certificate: 2379424

Master Contract: 249383

Project: 2439641

Date Issued: July 26, 2011

Issued to: CorDEX Instruments Ltd
1 Owens Rd
Skippers Lane Industrial Estate
Middlesbrough
Cleveland, TS6 6HE
United Kingdom
Attention: Richard Kidd

*The products listed below are eligible to bear the
CSA Mark shown with adjacent indicator 'US'*



Ron Wildish

Issued by: Ron Wildish

PRODUCTS

CLASS 2258 82 - PROCESS CONTROL EQUIPMENT - For Hazardous Locations -
Certified to US Standards

Class I, Div. 1, Groups B, C and D:

Model LM3000XP Laser Distance Meter; battery operated using Battery Pack P/N CDX2341-008; Enclosure Type 3R; Temp. Code T4, Tamb = -20C to +50C.

APPLICABLE REQUIREMENTS

FMRC 3600 – 1998 - Electrical Equipment for Use in Hazardous (Classified) Locations, General Requirements

FMRC 3615 – 2006 - Explosionproof Electrical Equipment, General Requirements

FMRC 3810 – 2005 - Electrical and Electronic Test, Measuring, and Process Control Equipment

ANSI/NEMA 250 – 2003 - Enclosures for Electrical Equipment